



Government of India  
Ministry of Health and Family Welfare  
**CENTRAL LEPROSY TEACHING AND RESEARCH INSTITUTE**  
**CHENGALPATTU – 603 001**

Director ☎044-27426275

Officer i/c stores ☎044-27429987  
E-Mail: [oics.cltri@nic.in](mailto:oics.cltri@nic.in)

No.F.5-31/2017 CS Vol-III

Dated 20.11.2017

To

All reputed firms/Suppliers

Sir,

Sub: - Open quotation for supply of Incubator & Hot Air Oven along with necessary Accessories – Reg.

Ref: This Institute Quotation No.12/F5-31/2017 CS dated 20.11.2017.

\*\*\*\*\*

Please submit your lowest rates/estimate for supply of the following laboratory Instruments along with necessary Accessories to this Institute subject to the following terms & conditions is as under. Details of the equipments and accessories with specifications mentioned in Annexure-I.

Item No.	Item Name	Specification	Indenter	Required Qty	Rate per Qty (in Rs.)
Annexure-I is enclosed					

2. Submit/send your estimate/quotation along with exact specification and catalogue in a sealed cover super scribing as **“Quotation for the supply of Incubator & Hot Air Oven along with necessary Accessories – Quotation No.12/F5-31/2017 CS dated 20.11.2017”** latest by **15<sup>th</sup> December, 2017.** Quotation without catalogue (wherever applicable) and after the due date will not be entertained. Details available in this Institute website : <http://cltri.gov.in> and Central Govt Tender Portal : <http://tenders.gov.in>.

  
DIRECTOR 20/11/17

**TERMS AND CONDITONS:**

1. Rate should be quoted only for the items, which are available in stock and can be supplied immediately on receipt of order. In case of high precision instruments the firms should have a guarantee certificate for their satisfactory performance.
2. Minimum three brands should be quoted.
3. The goods on receipt will be inspected and payment will be made only for the accepted number after satisfaction of the equipment.
4. **This institute is exempted from payment of S.T/VAT on purchase made within the state vide G.C.P.2280 Revenue dated 30.4.1973.**
5. No insurance charges are payable as per the rules in the Government. As such, the firms before quoting should take into consideration all the risks in the transit and then furnish quotations. If any point is raised as regards insurance charges after orders and issued, the same will not be entertained and the firms thereafter should effect the supply at their own cost.
6. Rates should be quoted F.O.R Chengalpattu.
7. Delivery period as per our schedule.
8. The Director shall have the right of rejecting the quotations in whole or part without assigning any reason therefore.
9. Copy of MSDS wherever applicable should be provided with the supply.
10. **A certificate to the effect that the rates quoted are equal or lesser than the market price and below the MRP.**

**Copy to:-**

1. Computer Section  
(Through Asst. Dir. Epid & Stats)
2. Central Govt. Tender Portal
3. Office Notice Board
4. Office File

- for wide publicity in Institute website

**CENTRAL LEPROSY TEACHING AND RESEARCH INSTITUTE  
CHENGALPATTU – 603 001**

Director ☎044-27426275

Officer i/c stores ☎044-27429987  
E-Mail: [oics.cltri@nic.in](mailto:oics.cltri@nic.in)

No.F.5-31/2017 CS Vol-III

Dated 20.11.2017

**ANNEXURE – I**

Item No.	Item Name	Specification	Indenter	Required Qty	Rate per Qty (in Rs.)
1	Incubator along with necessary Accessories	Thermostate controlled to maintained 37oC	Biochemistry Laboratory	1 No.	
2	Hot air Oven along with necessary Accessories	600 x 600 x 700 mm / 250 Ltr. Double layered transparent Glass window in-built into the door. Automatic cut off of heater & blower when door opened. Non-contact type door switch. Digital PID temperature controller with timer, alarms and auto tuning. Incoloy sheathed heater. Over temperature safety protection & MCB. Aero dynamic internal design. Solid and plain bottom without electricals. Stainless steel wire mesh shelves. Shelf height adjustable in 25mm steps. Mineral wool insulation. Silicon gasket in door. Temp range: ambient +10°C to 250°C. Control accuracy: ±0.2°C. Uniformity: ±2.0°C or better.	Animal House Laboratory	1 No.	



*Shew Kumar*  
**DIRECTOR**  
20/11/17